



Mass Flow Controller
Certification

Customer: NULL
Customer P/N: NULL
Serial Number: C110112019
Moder Number: FC-770A

Specifications

Calibration Date:	5/12/2015	Calibration Factor:	0.599
Flow :	200	Zero Offset:	0.003V
Rate:	SCCM	Test Gas:	N2
Gas:	SiH4	Semi Gas:	NA

Set Point %	Set Flow	Actual Flow	Error(FS)
0%	0.000	0.120	0.060%
25%	50.000	49.700	-0.150%
50%	100.000	101.230	0.615%
75%	150.000	151.300	0.650%
100%	200.000	199.650	-0.175%

Comments:

The Product described in this certificate was inspected and functionally tested in accordance with the customer purchase order, drawings, and specifications, and has satisfied Flows Instruments specification, inspection and testing, unless otherwise stated. Copies of the required inspection and test reports of the product are on file and available upon request.

The equipment used to calibrate the above device is in current calibration, which is traceable to the National Institute of Standards and Technology (NIST) or other equivalent international measurement standard.

Flows Instruments Technology Co. Ltd